This is a self-archived version of an original article. This version may differ from the original in pagination and typographic details.

Author(s): Giatzoglou, Alexandros; Poomaradee, Tanapoom; Pohjalainen, Ilkka; Rinta-Antila, Sami; Moore, Iain; Walker, Philip M.; Marmugi, Luca; Renzoni, Ferruccio

Title: A facility for production and laser cooling of cesium isotopes and isomers

Year: 2018

Version: Accepted version (Final draft)

Copyright: © 2018 Elsevier B.V.

Rights: CC BY-NC-ND 4.0

Rights url: <https://creativecommons.org/licenses/by-nc-nd/4.0/>

Please cite the original version:

Giatzoglou, A., Poomaradee, T., Pohjalainen, I., Rinta-Antila, S., Moore, I., Walker, P. M., Marmugi, L., & Renzoni, F. (2018). A facility for production and laser cooling of cesium isotopes and isomers. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 908, 367-375.
<https://doi.org/10.1016/j.nima.2018.08.095>

Accepted Manuscript

A facility for production and laser cooling of cesium isotopes and isomers

Alexandros Giatzoglou, Tanapoom Poomaradee, Ilkka Pohjalainen,
Sami Rinta-Antila, Iain D. Moore, Philip M. Walker, Luca Marmugi,
Ferruccio Renzoni

PII: S0168-9002(18)31053-2
DOI: <https://doi.org/10.1016/j.nima.2018.08.095>
Reference: NIMA 61136

To appear in: *Nuclear Inst. and Methods in Physics Research, A*

Received date: 26 July 2018; Revised date: 23 August 2018; Accepted date: 27 August 2018

Please cite this article as: A. Giatzoglou, et al., A facility for production and laser cooling of cesium isotopes and isomers, *Nuclear Inst. and Methods in Physics Research, A* (2018), <https://doi.org/10.1016/j.nima.2018.08.095>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

A Facility for Production and Laser Cooling of Cesium Isotopes and Isomers

Alexandros Giatzoglou^a, Tanapoom Poomaradee^a, Ilkka Pohjalainen^b, Sami Rinta-Antila^b, Iain D. Moore^b, Philip M. Walker^c,
Luca Marmugi^a, Ferruccio Renzoni^a

^a*Department of Physics and Astronomy, University College London, Gower Street, London WC1E 6BT, United Kingdom*

^b*Department of Physics, University of Jyväskylä, Surfontie 9, 40014 Jyväskylä, Finland*

^c*Department of Physics, University of Surrey, Guildford GU2 7XH, United Kingdom*

Abstract

We report on the design, installation, and test of an experimental facility for the production of ultra-cold atomic isotopes and isomers of cesium. The setup covers a broad span of mass numbers and nuclear isomers, allowing one to directly compare chains of isotopes and isotope/isomer pairs. Cesium nuclei are produced by fission or fusion-evaporation reactions using primary proton beams from a 130 MeV cyclotron impinging upon a suitable target. The species of interest is ejected from the target in ionic form, electrostatically accelerated, mass separated, and routed to a science chamber. Here, ions are neutralized by implantation in a thin foil, and extracted by thermal diffusion. A neutral vapor at room temperature is thus formed and trapped in a magneto-optical trap. Real-time fluorescence imaging and destructive absorption imaging provide information on the number of trapped atoms, their density, and their temperature. Tests with a dedicated beam of $^{133}\text{Cs}^+$ ions at 30 keV energy confirm neutralization, evaporation, and laser cooling to 150 μK , with an average atomic density of 10^{10} cm^{-3} . Availability of cold and dense atomic samples of Cs isotopes and isomers opens new avenues for high-precision measurements of isotopic and isomeric shifts thereby gaining deeper insight into the nuclear structure, as well

Email address: l.marmugi@ucl.ac.uk (
Luca Marmugi)

as for sensitive measurements of isotopes' concentration ratios in trace quantities. The facility also constitutes the core for future experiments of many-body physics with nuclear isomers.

Keywords: Laser cooling, Ultra-cold nuclei, Isotopes and isomers.

1. Introduction

Laser spectroscopy of radioactive atoms has the potential to shed new light on open questions about nuclei and fundamental symmetries [1–5]. However, progress in this field faces many experimental challenges: in addition to the main problems constantly addressed in atomic and laser physics, the isotopes' short half-life requires dedicated experimental facilities and protocols, at the boundary between nuclear and atomic physics.

In this context, cold and ultra-cold atoms offer a unique opportunity, allowing one to collect large samples of rare and unstable isotopes, and providing the conditions for high precision laser spectroscopy – with sub-Doppler linewidths of the order of a few MHz – and long interrogation times – many orders of magnitude longer than the lifetime of the electronic state excited by lasers [6, 7]. For example, laser cooling of heavy unstable atoms (mostly Fr [8, 9]) has given access to unprecedented spectroscopic data, motivated by the study of phenomena such as parity non-conservation in atoms and the permanent electronic dipole moment [7, 10–12].

In this paper, we report on the facility for laser cooling of cesium isotopes and isomers built at the Accelerator Laboratory of the University of Jyväskylä (Finland). From the production site of the ions of interest to the trap of cold atoms, the energy scale spans from 10^4 eV (10^6 eV in the case of the primary proton beam) to 10^{-8} eV in around 5 s. The facility, the first of its kind entirely dedicated to Cs, is designed to trap wide mass range of Cs isotopes (from $A=133$ to 142), in their nuclear ground and long-lived excited states ($A=134m$, $135m$, $136m$, $138m$). This opens up new avenues for the investigation of charge radii anomalies in $^{A,\text{Am}}\text{Cs}$ pairs [13, 14], and for the creation and validation of nuclear

forensics techniques and instrumentation [15, 16]. The ability to manipulate ultra-cold radioactive atoms will also allow the exploration of yet-to-be observed many-body effects [17].

The setup can be ideally divided into a high-energy section (production, acceleration, and mass separation of ions) and a low-energy section (thermalization, neutralization, and laser cooling and trapping). The description of the experimental facility presented here follows the order of operation: in Sec. 2 an overview of the facility is presented. In Sec. 3 a detailed description of the high energy section of the experimental setup is provided, while in Sec. 4 the low energy (“cold atoms”) section is described. Finally, in Sec. 5, a full experimental characterization of the facility with a $^{133}\text{Cs}^+$ beam is reported. Results obtained during implantation tests with $^{135\text{m},135}\text{Cs}^+$ are also reported.

Figure 1: (Color online) Overview of the IGISOL facility. The vacuum chamber of the cold atoms experiment is visible on the right hand side. The optical and laser set-up is not shown for clarity. Details are in the text.

2. Overview of the Facility

An overview of the experimental setup, comprising the Ion Guide Isotope Separation On-Line (IGISOL) mass separator, an off-line source of $^{133}\text{Cs}^+$, and the low energy section, where laser cooling is performed, is presented in Fig. 1.

The primary beam is produced in the K130 cyclotron of the Accelerator Laboratory of the University of Jyväskylä. Fission (e.g. in the case of $^{135,135m}\text{Cs}$, discussed in Sec. 5.2) or fusion-evaporation (e.g. in the case of $^{134,134m}\text{Cs}$, promising for radius anomaly studies [18]) products following nuclear reactions are electrostatically accelerated towards the low-energy sector. Here, a dedicated γ spectroscopy line – within the IGISOL facility – can be used for monitoring the isomeric content of the beam. Alternatively, the ionic beam is routed towards the science chamber, where ions are neutralized via thin foil implantation, extracted as an atomic vapor, and trapped in a Magneto-Optical Trap (MOT). The atomic sample – cooled down to 150 μK – is monitored in real-time with fluorescence imaging, or with absorption imaging with a repetition rate of 0.1 Hz. For testing the operation of the system, and retaining the possibility of direct comparison with the stable reference isotope, an off-line source of $^{133}\text{Cs}^+$ was realized on the second floor of the facility (top of Fig. 1). The ions from the off-line source, once merged into the main separator, share the same path with the radioactive beam. This also allows testing and optimisation of the transport line.

3. High-Energy Section

At IGISOL, nuclear reactions are induced using either light ions (protons, deuterons or alpha particles) or heavy ions from either a new MCC30 light-ion cyclotron, or a K130 cyclotron [19]. In this work, the K130 cyclotron is used, with a 50 MeV proton beam impinging on a natural U target in order to induce fission. The fission products recoil isotropically from the target into a buffer gas cell located in the IGISOL target area (Fig. 1) and operating with helium at typical pressures of 300 mbar. Approximately 1% of the fragments are thermalized and stopped in the gas [20], with the majority of isotopes surviving as singly-charged ions. The reaction products, along with the buffer gas and any impurities, are swept out of the gas cell through a 1.2 mm diameter exit hole, whereby the neutral gas is pumped away using high-throughput vacuum

pumps.

The ions are subsequently captured in the radiofrequency field of a sextupole ion guide (SPIG) [21] and are guided towards the high vacuum region of the mass separator. At this stage, the ions are accelerated to a potential of 30 keV, electrostatically transported, and mass separated using a 55° dipole magnet with a typical mass resolving power $300 \leq M/\Delta M \leq 500$.

3.1. Off-line Source of $^{133}\text{Cs}^+$

Rather recently, a new off-line ion source was constructed and installed in order to provide stable beams of ions from a location on the second floor of the IGISOL facility (Fig. 2).

Figure 2: (Color online) Stable ion source (“off-line source”). A surface ion source with the desired species – in the present case ^{133}Cs – is inserted on the left hand side. Ionization of Cs and other alkali elements occurs via the hot surface of a filament. The ions are then electrostatically accelerated by a 30 kV electrode and are routed to the IGISOL mass separator.

The use of the off-line source allows for a more flexible and safe operation, with off-line operation available immediately following an on-line experiment without the need to wait for a “cooling” period. It also enables a stable reference source to be provided in-situ with radioactive beams from the front-end of the

85 facility. Two ion sources are in current operation, a surface ion source and a so-called discharge ion source. The latter has been presented recently in connection with the production of an ion beam of stable ^{89}Y for collinear laser spectroscopy [22].

The surface ion source is used for the production of a $^{133}\text{Cs}^+$ beam. Ions
90 are accelerated using a dedicated high voltage platform to the same operating potential as the front-end of the IGISOL, namely 30 kV. Such a beam of ions is mass separated in an identical manner to the on-line operation and transported to the science chamber of the low-energy section for further manipulation.

4. Low-Energy Section

95 The low-energy section is located immediately downstream from the focal plane of the IGISOL mass separator, occupying a new beam line, highlighted in Fig. 1. This section is differentially pumped, and isolated by an ultra-high vacuum (UHV) pneumatic gate valve after the IGISOL switchyard. A 350 l s^{-1} turbo-molecular pump (Leybold CE TURBOVAC 350i), ensures a stable vac-
100 uum $\leq 10^{-8}$ mbar, even when the gate valve is abruptly opened. A custom-made CF63 cross element secures the turbo-molecular pump to the main transport line, provides support for the science chamber, and allocates diagnostic tools such a movable Faraday cup, and a Penning gauge. The cross element is connected to a CF63/CF40 adaptor, which in turn contains a second vacuum
105 impedance.

As shown in Fig. 3, an ion pump (20 l s^{-1} , Varian VacIon Plus 20 StarCell) produces a residual pressure in normal operation of 1×10^{-9} mbar. The ion pump is isolated from the experimental chamber during maintenance by a second UHV gate valve. We have observed that – together with venting with N_2
110 – this dramatically reduces the waiting time for breaking and re-establishing operational vacuum conditions. A few hours, as opposed to two days, are usually enough, allowing emergency maintenance to be performed. The ion pump’s residual magnetic field at the beam’s position has been taken into account during

transport optimization. Accordingly, no mu-metal shielding is used.

115 A CF40 cross provides connection to a CF40 valve, used for pre-evacuation
with a removable 50 l s^{-1} turbo-molecular pump and for N_2 venting. On the
opposite side, a CF16 valve is installed and connected to a blind CF16 flange,
where a reservoir of ^{133}Cs has been sealed. Atoms are extracted – when needed
– via thermal evaporation, and are usually detected in the main experimental
120 chamber within a few seconds after the valve has been opened. Cs vapor pro-
vides another tool for testing the optical setup and the operation of the MOT.
Attached to the remaining port of the CF40 cross is the experimental chamber,
whose main axis coincides with that of the ion transport line.

4.1. Science Chamber

Figure 3: (Color online) Rendering - to scale - of the terminal part of the low energy section. The borosilicate glass science chamber is shown on the right hand side. The ion beam enters from the CF40 connection on the left, next to the 20 l s^{-1} ion pump, passes through the gate valve, and enters the laser cooling chamber. A CF16 blind flange equipped with electrical feedthroughs supports an implantation foil. Anti-Helmholtz quadrupole coils are visible at either side of the glass chamber. The implantation thin foil (not visible) is placed 20 mm inside the science chamber. Further details are shown in Fig. 4. Positions of the laser beams are marked by red cylinders. A solid reservoir of ^{133}Cs is sealed by the CF16 valve visible in the background.

125 The science chamber is a borosilicate glass spheroid, equipped with 40 mm
diameter, 40-20 scratch-dig optical windows, supported by dedicated borosili-
cate glass tubes, and enclosed by two coils (Fig. 3). The chamber inner diameter
is around 100 mm. The overall length is 250 mm, including the connection to
the vacuum line and the connection to the neutralizer feedthrough as described
130 below. Optical windows are aligned in pairs to allow easy alignment of the laser
beams for the MOT. Two are placed along the axis orthogonal to the transport
line, along the main axis of the main coils. The other two pairs of optical win-
dows are located in the plane parallel to the coils' faces with a separation of
around 80°.

135 Connection to the vacuum line is ensured by a CF40 metal-to-glass con-
nector. This also acts as a mechanical support for the whole chamber. Extra
support is provided by tailored foam segments underneath the two “horizon-
tal” windows resting on the main coils mounts (not shown in Fig. 3). This
also dampens mechanical vibrations, without producing sheer stress that could
140 damage the chamber in the long term.

The two main coils are in the ideal anti-Helmholtz configuration, generat-
ing a magnetic quadrupole field. The system is air-cooled and creates a linear
magnetic field gradient of 10 G/cm/A in the trapping region. Current in the
coils is controlled by a water-cooled MOSFET (IXFN 120N20), driven by an
145 SRS SIM960 PID controller. The setpoint of the PID loop is controlled via Lab-
VIEW, and can be changed during the experimental sequence via the control
software [23]. In addition, three pairs of mutually orthogonal squared compen-
sation coils (not shown in Fig. 3) nullify the background DC fields acting on the
atoms, and control the position of the atomic trap.

150 4.1.1. *Anti-relaxation Organic Coating*

The inner surface of the experimental chamber is coated with a layer of poly-
dimethylsiloxane (PDMS, $\text{CH}_3[\text{Si}(\text{CH}_3)_2\text{O}]_n\text{Si}(\text{CH}_3)_3$). The coating is a strat-
egy for minimizing atomic losses for high efficiency laser cooling experimental
chambers [24–26]. In detail, PDMS lowers the glass surface adsorption energy

155 down to ~ 0.1 eV, drastically reducing chemisorption after atom/wall collisions. Consequently, the amount of atoms of interest available in the experimental chamber is increased. This is a paramount feature in the case of short-lived species, which allows the cold atom MOTs to be maximally populated.

The chosen PDMS (Sigma-Aldrich) has a viscosity of 65 cSt. For deposition, we have adopted a protocol based on a modified version of the procedure presented in [26]. The glass is prepared in a prolonged bath of 0.45:0.45:0.10 volume solution of CH_3OH , $\text{C}_2\text{H}_5\text{OH}$, and KOH . KOH is slowly and carefully dissolved in the liquid solution with gentle stirring, making sure that the mixture's temperature does not exceed 40°C . The chamber is rinsed five times
165 with de-ionized water, and then is left in a ventilated chemical hood at room temperature for 30 min. The chamber is then connected to a turbo-molecular pump (50 l s^{-1}). Once a stable pressure of 1×10^{-6} mbar is reached, gentle baking with resistive heating tape is performed. Three layers of Al foil isolate the system from air and ensure uniform heating of the chamber. The temperature
170 is increased at a rate of $10^\circ\text{C}/20$ min, until 120°C . The procedure is terminated when the pressure of 1×10^{-7} mbar is stable. Cooling of the chamber is obtained with spontaneous thermalization, after switching off the heating tape.

The coating is applied by spin-coating (0.1 rps) of a solution 0.05:0.95 of PDMS and $(\text{C}_2\text{H}_5)_2\text{O}$. After five minutes, the spin is interrupted and the excess
175 solution is eliminated. The chamber is then baked up to 100°C , at a rate of $10^\circ\text{C}/20$ min, until the baseline pressure of 1×10^{-7} mbar is reached. After installation in the main vacuum line, baking of the whole system is performed. A baseline pressure of 1×10^{-9} mbar is usually obtained within 24 h. A final curing ("passivation") is obtained by using the thermal Cs vapor from the solid
180 Cs reservoir. The progression of the curing process is monitored by observing the stability of the number of atoms trapped in the MOT obtained from thermal vapor.

4.2. Ion Neutralization and Atom Extraction

The incoming ion beam is slowed down and neutralized via thin foil im-
 185 plantation [15, 27]. Ions are electrostatically focused on a 25 μm thick yttrium
 foil (16 mm $\times$ 9 mm), held approximately 20 mm inside the science chamber (for
 details and the actual position, see Figs. 4).

Figure 4: (Color online) Sketch - to scale - of the neutralizer mount. A thin Y foil, where
 the incoming ion beam is focused, is supported by copper contacts, in turn connected to
 electrical feedthroughs. Mechanical stability is ensured by ceramic rods, secured to a CF16
 flange. Resistive heating allows one to control the temperature of the foil, up to 10^3 K. Top:
 relative position of the Y foil and the MOT (approximately at the center of the laser beams
 overlapping region). Bottom: detailed view of the neutralizer mount.

The Y foil (99.9% REO, supplied by Alfa Aesar) is kept in place by copper
 supports, connected to two hollow aluminum oxide (Al_2O_3) rods, 105.6 mm long.
 190 The rods are mechanically secured to the 20 A-rated electrical feedthroughs of a
 CF16 flange. Electrical connections are ensured by a hollow Cu core embedded
 in the ceramic rods, connecting 1.5 mm diameter Cu wire of the feedthroughs
 to the supports of the Y foil. With a newly installed foil, the circuit has a
 resistance of $1.2 \pm 0.1 \Omega$, where the uncertainty takes into account the statistical
 195 variability observed. A MACOR centering ring, of 16.5 mm diameter, ensures
 the correct positioning of the supporting rods and prevents excess tension on

the thin Y foil. At the same time, mechanical tolerance allows free expansion of the foil upon increase of the temperature.

SRIM simulations [28] indicate a peak implantation density around 150 Å from the surface of the foil for the 30 keV ions (Fig. 5). With respect to recent works with Fr [26], the implantation depth is significantly larger, causing a greater involvement of the bulk of the foil in the storage of the incoming ions. This configuration is more efficient in the case of long-lived particles, such as those of interest here.

Figure 5: (Color online) SRIM simulations of the implantation of 30 keV $^{133}\text{Cs}^+$ in the 25 μm thick Y foil: linear density of the implanted ions. For this simulation, 10^4 ions were used, with a target length of 100 nm.

Implanted ions dissipate a large fraction of their kinetic energy via inelastic collisions with the metal lattice, and slow down to a diffusive regime. Because of the favorable imbalance between the Cs ionization potential ($I_{\text{Cs}}=3.89 \text{ eV}$ [29]) and the Y work function ($W_{\text{Y}}=3.1 \text{ eV}$, as estimated for a pure Y sample), they also acquire an extra electron from the surrounding metal, thus achieving neutralization required for laser cooling. Inside the foil, the relative concentration of neutral atoms (n) and ions (n_+) is given by the Langmuir-Saha equation [30]:

$$\frac{n_+}{n} = \frac{g_+}{g} \exp\left(\frac{W_Y - I_{Cs}}{k_B T}\right), \quad (1)$$

where k_B is the Boltzmann constant, T is the foil's temperature in K, and $g_+/g = 1/2$. In operational conditions, $n_+/n \leq 10^{-4}$. The value of W_Y used in Eq. 1 may also vary with the surface's lattice structure and contamination.

215 We note that an alternative candidate for the neutralizer foil is Zr, as recently adopted in Fr experiments with promising results [31].

Atoms are extracted by thermal diffusion, enhanced by heating the foil up to 10^3 K, via controlled flow of a DC current (I_{DC}). An optical pyrometer (METIS M3 SensorTherm) is used for measuring the characteristic curves T_Y - I_{DC} . We
220 have observed a large variability in the Y foil response with a given DC power supplied, up to a factor 2. We attribute this to the intrinsic purity of the sample: although quantitative analysis was not possible, the nominally most pure foils require a larger current for reaching a given temperature. This is consistent with the presence of fewer impurities that act as scattering centres for the flowing
225 charge carriers. To reduce the impact of contamination during installation, the Y foil is kept in the original Ar atmosphere as long as possible. It is then extracted in air and installed on the dedicated mount in a clean environment at atmospheric pressure. Afterwards, the foil is immediately installed in the science chamber, which has been in the meantime vented with N_2 . Evacuation is
230 immediately started. As soon as stable vacuum is achieved, progressive heating is performed to enhance controlled outgassing of the freshly installed Y foil. The actual rate of heating may vary depending on the vacuum conditions, which are constantly monitored.

Upon thermal desorption in the science chamber, the atoms' velocity distribution is determined by the high temperature required for efficient desorption
235 ($600 \text{ K} \leq T \leq 1000 \text{ K}$). Atoms released in this way are too fast to be efficiently laser cooled, given the finite capture range of the cooling mechanism. The PDMS coating allows atoms to thermalize at room temperature: atoms are desorbed with a thermal velocity distribution compatible with room temperature.

240 In this way, the percentage of atoms with velocities below the MOT capture velocity is increased [24].

4.3. ^{133}Cs MOT Configuration and Laser System

The MOT is formed by three retro-reflected optical beams, to increase the compactness. Cooling and repumper beams are overlapped and then split in
245 three components, each controlled in power by dedicated polarizing beamsplitter cubes and half-waveplates. Beam-expanders set the size of the laser beams to 20 mm (diameter). Circular polarization is independently obtained with dedicated quarter-waveplates, as standard in cold atom experiments.

The cooling laser light is generated by an extended-cavity diode laser (ECDL) equipped with an intra-cavity interference filter (Radiant Dyes NarrowDiode)
250 tuned to the $6^2S_{1/2}(F=4) \rightarrow 6^2P_{3/2}(F'=4 \times 5)$ cross-over transition of the D_2 line. A chain of a double-pass (AOM1) and a single-pass (AOM2) acousto-optic modulators shifts the laser frequency to the $6^2S_{1/2}(F=4) \rightarrow 6^2P_{3/2}(F'=5)$ transition. AOM1 allows the operator to control the fine tuning of the laser,
255 while AOM2 provides fast switching of the laser beam during the experimental sequence [23]. The cooling laser light is amplified by a tapered amplifier (m2k GmbH TA-0850-2000). A single-mode polarization-maintaining optical fiber acts as a spatial filter, ensuring a TEM_{00} transverse mode. The intensity of each cooling beam in the trapping region is 4.67 mW cm^{-2} , corresponding
260 to around $4.2I_{\text{sat}}$, where $I_{\text{sat}}=1.10 \text{ mW cm}^{-2}$ is the saturation intensity of the cycling transition used for laser cooling.

The repumper laser light is supplied by a second ECDL (Radiant Dyes NarrowDiode), tuned to the $6^2S_{1/2}(F=3) \rightarrow 6^2P_{3/2}(F'=4)$ optical transition. This laser compensates for hyperfine optical pumping losses, as typical in a
265 MOT. The average repumper intensity is 1.1 mW cm^{-2} . A mechanical shutter controls the beam during the experimental sequence.

Laser frequency locking is obtained via Dichroic Atoms Vapor Laser Lock (DAVLL) [32], and custom designed PI controllers. DAVLL provides locking stability in a noisy environment for several hours, independent of possible laser

270 power fluctuations.

4.4. Cold Atoms Imaging and Diagnostics

The MOT is monitored with two different approaches, fluorescence and absorption imaging. Absorption imaging relies on a resonant beam ($6^2S_{1/2}(F = 4) \rightarrow 6^2P_{3/2}(F' = 5)$, $50 \mu\text{W}$), intercepting the MOT along the main coils' axis. This beam is independently controlled by an additional AOM (AOM3), driven by the experimental LabVIEW routine. After interaction with the atomic cloud, the beam is imaged by a computer-triggered CCD camera (PCO PixelFly). Absorption imaging – destructive for the trap – is chosen for measuring the MOT's size, density, temperature, and lifetime. Temperature measurements are obtained with free-expansion and time-of-flight measurements [23].

Fluorescence imaging collects the spontaneous emission from the trapped atoms during laser cooling. A CCD camera (PCO PixelFly), equipped with an interference filter ($\lambda_c = 852 \text{ nm}$, full-width at half maximum FWHM=10 nm), images the MOT. A LabVIEW programme continuously monitors the MOT population, its distribution in space, and the cooling laser power stability. The software automatically subtracts the scene's background and compensates for laser intensity fluctuations. An ultimate background noise corresponding to ~ 20 atoms, with a residual fluctuation of only 3% in 1 h, has been measured. This number comprises optical, thermal, and shot noise. Fluorescence imaging can monitor the “real-time” evolution of the MOT population, and is therefore well-suited for investigation of the trap dynamics, of the contribution of the ionic beam, and for tuning and optimization [8, 26].

5. Experimental Characterization and Commissioning of the Facility

The operation of the facility is characterized in the following. In particular, the first demonstration of MOT real-time loading from $^{133}\text{Cs}^+$ is presented as a validation test for the transport and neutralization setup. The relatively high currents achieved with the beam of stable $^{133}\text{Cs}^+$ ($\sim 10^{-9} \text{ A}$) cannot be representative of the low currents achievable with radioactive beams. Nevertheless,

this is beyond the scope of the present work. Furthermore, scaling with current
 300 can give a first approximated estimation of the system's response with a smaller
 number of atoms.

5.1. $^{133}\text{Cs}^+$ -loaded MOT

In the science chamber, ^{133}Cs MOTs can be obtained from different sources.
 One method is direct loading from thermal vapor produced by the neutral ce-
 305 sium sample contained in the CF16 reservoir. This corresponds to the “conven-
 tional” loading process for a MOT.

In operational conditions, the residual background partial pressure of Cs
 vapor is negligible, and no MOT loading can be observed. This condition corre-
 sponds to that of radioactive species. In this case, loading can be obtained only
 310 via desorption from the Y foil and thermalization on the PDMS coating either
 from previously implanted neutralized ions, or from freshly implanted neutral-
 ized ions. Figure 6 shows a fluorescence image – in false color – of a ^{133}Cs MOT,
 together with the real time MOT profiles, calculated by the control software.
 The MOT in Fig. 6 is obtained with loading from the Y foil, after implantation
 315 with $^{133}\text{Cs}^+$, at an average current of 1×10^{-10} A.

No fundamental difference was found among MOTs loaded with different
 methods. We attribute this to the effectiveness of the neutralization and ex-
 traction processes conveyed by the Y foil and the PDMS coating. On aver-
 age, the optimum cooling laser detuning was found to be $\Delta = -2.65\Gamma$, where
 320 $\Gamma = 2\pi \times 5.2\text{MHz}$ is the $F = 4 \rightarrow F' = 5$ transition's natural linewidth. The
 typical MOT density is 10^{10}cm^{-3} , with a temperature $T=(150 \pm 10)\text{ }\mu\text{K}$. The
 uncertainty takes into account instrumental uncertainties and statistical fluctua-
 tions. A very small asymmetry ($\pm 1\text{ }\mu\text{K}$) was measured along the two orthogonal
 axes, because of a very good expansion symmetry. This confirms effective com-
 325 pensation of the background fields and proper alignment of the imaging system.
 We note that – as expected – the MOT temperature can significantly vary,
 depending on the chosen parameters.

Figure 7 shows N_{MOT} as a function of time, when I_{DC} is increased by 0.5 A

Figure 6: (Color online) Typical fluorescence image of a ^{133}Cs MOT, loaded by heating the Y foil to 931 K ($I_{\text{DC}}=5.5$ A). The trap population is $N_{\text{MOT}}=1700$. The foil has been previously implanted with 1×10^{-10} A of 30 keV $^{133}\text{Cs}^+$. The MOT is therefore loaded in the absence of the $^{133}\text{Cs}^+$ beam. Insets: real time MOT profiles (rows and columns pixel sums). CCD camera exposure is 2 s.

between 3.5 A (691 K) and 6.5 A (1010 K) every 100 s. The black thin trace shows
 330 the behavior when – before the described sequence – the Y at room temperature
 is exposed to a Cs thermal vapor 10^{10} cm^{-3} for 20 min. In this case, atoms are
 adsorbed at the surface sites of the Y foil, and therefore are quickly ejected upon
 heating. This creates a spike in the MOT population, which is however rapidly
 335 reduced, given the finite density of superficially adsorbed atoms. Such rapid
 dynamics are typical of atoms adsorbed in surface sites [30, 33]. Noticeably,
 the N_{MOT} level reached after 50 s is higher than the initial level, thus indirectly
 confirming the quality of the PDMS coating: atoms can interact several times
 with the coated walls before being permanently lost because of adsorption or
 the vacuum system. The progressive decrease of the maximum peak at higher
 340 temperatures indicates that the main surface source of atoms is rapidly depleted.
 The behavior dramatically changes at high currents. In particular, at 1010 K
 ($I_{\text{DC}}=6.5$ A), a net increase of the atomic population and completely different

loading dynamics are observed. The slow process is typical of bulk processes [33].

Figure 7: (Color online) Different loading methods via Y. Thin black line: after 20 min exposure to 10^{10} cm^{-3} (multiplied by 0.25). Red squares: Previously implanted beam. Blue disks: extra 10 min of $^{133}\text{Cs}^+$ at $1 \times 10^{-10} \text{ A}$ at the Y foil. The vertical dashed lines mark increase in T_Y . From left to right: $T_Y = 691, 767, 832, 887, 931, 972, 1010 \text{ K}$.

When the vapor contribution to the Y is completely eliminated (red squares in Fig. 7), the fast dynamics completely disappear. At high temperature, the bulk contribution due to previous implantations emerge again. In this condition, substantial MOT loading – only with slow dynamics – is observed only at high temperatures ($T_Y=972 \text{ K}$, 1010 K , corresponding to $I_{DC}=6.0 \text{ A}$, 6.5 A).

The same experiment is performed after further implantation of “fresh” $^{133}\text{Cs}^+$ ions, at an average current of 10^{-10} A . Remarkably, MOT loading can be clearly observed already with $I_{DC}=4.5 \text{ A}$, and the fast dynamics is again observed. At higher current, the dynamics progressively approach the slow regime attributed to the bulk contribution, although almost two times bigger than what was observed before the extra implantation. This is a direct evidence of the contribution – both on the surface and bulk populations of the Y foil – of the $^{133}\text{Cs}^+$ ionic beam to the MOT population.

5.1.1. Loading from previous ion implantations

Figure 8 shows results concerning ^{133}Cs MOT loading from previously im-
 360 planted ions. It is noteworthy that in all cases there is no detected loading from
 background vapor.

Figure 8: (Color online) MOT loading from Cs atoms desorbed from Y, after previous $^{133}\text{Cs}^+$ implantations. N_{MOT} vs T_Y . Dashed lines marks the background noise level, the dotted lines the 3σ level. The yellow line is an exponential fit based on the Arrhenius equation applied to desorption. Inset: detail of the low-current regime, where values are consistent with no detection.

In this regime, desorption of neutral atoms from Y foil is produced only by thermal diffusion: previously implanted ions, once neutralized inside the foil, progressively diffuse out of the foil, with a clear exponential dependence on T_Y .
 365 This behavior is consistent with a purely thermal diffusion process.

To confirm the gain in N_{MOT} produced by the ion beam, we investigated the impact of previously implanted $^{133}\text{Cs}^+$ in Fig. 9. Here, 1×10^{-10} A of $^{133}\text{Cs}^+$ are implanted for a variable time t_{ion} , and the maximum population of the MOT is plotted against the implantation time ($T_Y=930$ K). The data confirms the
 370 expected growth in the number of trapped atoms for increasing implantation time.

Figure 9: (Color online) Loading exclusively from Cs atoms desorbed from Y, as a function of implantation duration ($t_{\text{implantation}}$) with 10^{-10} A. Y at 5.5 A. Quadratic fit is a guide for the eye. The dashed line marks the background noise level. Atoms trapped at $t_{\text{implantation}} = 0$ min were previously implanted. This level can be considered an offset when evaluating the impact of $t_{\text{implantation}}$.

5.1.2. Direct loading from neutralized $^{133}\text{Cs}^+$ beam

A second set of validation experiments was carried out to find the direct evidence of MOT loading from freshly implanted ions. In particular, the compatibility of the 30 keV beam and the MOT is also assessed. This is an essential prerequisite for continuous implantation operation with radioactive beams.

To this goal, a background MOT is loaded from previous implantations, by heating the Y foil. Once the population is stable, the $^{133}\text{Cs}^+$ beam is switched on for 300 s, and then switched off again. A typical result is shown in the inset of Fig. 10. Here, we note a significant increase of N_{MOT} both with initial zero population (yellow trace) and with initial substantial population (green trace). The latter also confirms the compatibility of the MOT and the ion beam. In both cases, the response of the MOT population to the incoming beam is observed in around 5 s after the ion beam is switched on. Noticeably, after the beam is

Figure 10: (Color online) Direct loading from neutralized $^{133}\text{Cs}^+$ beam: N_{MOT} before (blue disks) and after (red diamonds) exposure to 10^{-10} A $^{133}\text{Cs}^+$ beam, as a function of T_Y . Inset: MOT fluorescence measured at $T_Y = 691$ K and 887 K ($I_{\text{DC}} = 3.5$ A and 5 A). The vertical dashed lines mark the timestamps of magnetic field gradient on ($t=-30$ s), ion beam on ($t=0$ s), ion beam off ($t=300$ s), magnetic field gradient off ($t=400$ s). The horizontal dashed lines mark the detection threshold.

switched off, the MOT population decays to a larger level than the initial one.

There is a relatively small and constant increase in N_{MOT} when the ion beam is on for $T_Y \leq 767 \text{ K}$ ($I_{\text{DC}} \leq 4 \text{ A}$). On the contrary, from $T_Y \geq 832 \text{ K}$ ($I_{\text{DC}} \geq 4.5 \text{ A}$), a substantial increase of the MOT population is observed, which increases with the foil temperature. This is consistent with thermally activated diffusion of neutralized atoms. The role of thermal diffusion is further confirmed by the larger increase in the MOT population observed in the inset of Fig. 10 with increasing T_Y : this effect is attributed to more effective desorption of ions implanted in deeper regions of the foil, aided by the higher temperature.

This regime corresponds to the most direct evidence of operation of the complex ion beam/neutralization/organic coating/laser trapping. Therefore, it is also tested when no loading from vapor or the Y foil is observed (i.e. when there is no MOT unless new ions are implanted and neutralized). This mimics the working conditions with radioactive atoms, where the instability of the species of interest can make “off-line” loading practically impossible. In Fig. 11, such a regime is demonstrated and characterized. Evidence of direct loading from the ion beam (on at $t=0 \text{ s}$ and off at $t=300 \text{ s}$) is observed at any temperature between 607 K and 972 K (I_{DC} between 3.0 A to 6.0 A). A temperature-independent increase of N_{MOT} is also observed at $T_Y \leq 767 \text{ K}$ ($I_{\text{DC}} \leq 4 \text{ A}$). We also observed this effect well below 400 K , without any obvious thresholds in temperature.

Such an increase at low T_Y – independent from the Y heater current and also observed in Fig. 10 – can be explained with direct ejection of previously neutralized atoms at the surface of the foil by the high-energy incoming ions. This effect, practically similar to sputtering, can be directly observed here because of the high kinetic energy of the incoming ions (30 keV).

With the currently available diagnostic on the ion beam in the science chamber (see also Sec. 6), we did not observe any increase in MOT losses due to the presence of the incoming ions. This can be also attributed to the small vertical offset between the MOT center and the ion beam core, as determined by optimizing the position of the trap for maximum population.

Figure 11: (Color online) Direct contribution from the Y foil. Maxima in N_{MOT} versus Y temperature. 10^{-10} A of $^{133}\text{Cs}^+$. Note effect of sputtering at $T_Y \leq 767$ K. Inset: real-time fluorescence evolution at different Y temperatures (from bottom to top: $T_Y = 649, 691, 767, 831, 887, 929$ K).

Notwithstanding the currently limited direct information on the number of incoming ions and on the “freshly” desorbed ^{133}Cs atoms (due to the bulk contribution), a lower bound for the trapping efficiency based on the results shown in Fig. 11 can be estimated. By using 6×10^8 atoms/s (i.e. 10^{-10} A of $^{133}\text{Cs}^+$), and an average MOT lifetime of $\tau = 2.85$ s (1/e time, measured in a separate experiment), the overall efficiency at $T_Y = 929$ K is: $\eta \geq 3 \times 10^{-6}$.

5.2. Implantation test runs

Lastly, we report on the implantation test runs conducted with a $^{135,135m}\text{Cs}^+$ beam. This allowed us to validate the production, transport, mass selection, and implantation procedure with one of the species of interest [17]. ^{135m}Cs ($J^\pi = 19/2^-$) has a half-life of $T_{1/2} = 53$ min, before decaying via an M4 ($E_\gamma = 846.1$ keV) and an E2 ($E_\gamma = 786.8$ keV) γ transition to the ground state ^{135}Cs ($J^\pi = 7/2^+$, $T_{1/2} = 2.3 \times 10^6$ yrs). Because of these characteristics, it is well-suited for implantation tests: it can be directly detected in the ion beam,

Figure 12: Beam spectra with mass number $A=135$, after proton-induced fission of a U target and extraction. Left: γ spectrum produced by a Ge detector after 30 min implantation. ^{135m}Cs implantation rate is $1.5 \times 10^4 \text{ s}^{-1}$. Right: Off-line decay curve of both γ transitions of ^{135m}Cs , obtained after implantation in the Y foil. Inset: γ spectra of the implanted foil at selected timestamps after implantation. The ^{135m}Cs γ peaks are highlighted by black fitting curves.

430 or it can be monitored after implantation to observe the γ decay.

During this test, both the ground- and the isomeric state of ^{135}Cs are populated by proton-induced fission of natural U using a primary beam intensity up to 6 μA . To verify the production of the isomeric state, the mass separated $A/q=135$ beam is routed towards the spectroscopy line located to the right of the electrostatic switchyard. The ion beam is then implanted for 30 min into a thin foil mounted at the end of the beam line which is directly viewed from behind by a Ge detector (Canberra GC7020, energy range from 40 keV to 10 MeV) in order to detect the emitted γ photons. In Fig. 12, we show the “on-line” γ spectra measured with the Ge detector. Clear signatures of the M4 and E2 γ emissions are observed. In addition to detection during implantation, γ emissions are also monitored during a 30 min decay period when no beam is implanted. Following analysis of the 787 keV and 846 keV peaks – taking into account the solid angle and detector efficiency – an implantation rate of $2 \times 10^4 \text{ s}^{-1}$ was detected, with 5 μA intensity of the primary beam.

445 A transmission efficiency of $\sim 70\%$ has been measured from the switchyard to the end of the spectroscopy line via implantation of $A/q=112$ fission products into foils mounted directly in front of Si charged-particle detectors. In this case, ^{112}Rh ($T_{1/2} = 6.8 \text{ s}$) is used for beam tuning due to the suitable half-life. To verify the corresponding value from the switchyard to the neutralizer foil in the science chamber, a trace amount of stable ^{136}Xe gas is added to the He buffer gas. The Xe is ionized during on-line operation of the gas cell and can be detected in quantities suitable for Faraday cups. During the on-line production of ^{135m}Cs , 95 nA of ^{136}Xe was detected in the switchyard. Of this, 38 nA was measured at the Y neutralizer foil, thus leading to a transmission efficiency of $\sim 40\%$. On the basis of these numbers, it is therefore estimated that $1.2 \times 10^4 \text{ s}^{-1}$ ^{135m}Cs were implanted into the neutralizer foil.

450 ^{135}Xe γ signatures are also detected (Fig. 12a)), as well as those of other by-products of the fission reaction. We note that this will not represent a problem for the purity of the cold atomic samples: further selection – beyond the charge-to-mass ratio 135 – will be automatically performed by the laser cooling process.

Unwanted species (or ground state, as opposed to the isomeric state) will not be trapped.

To confirm successful implantation of ^{135m}Cs , in Fig. 12, we show the “off-line” γ spectra, measured by extracting the Y foil from the vacuum chamber after 2 d implantation at 3×10^{-14} A $^{135m}\text{Cs}^+$. Data are obtained in a low-
 465 background off-line chamber, at regular time intervals, starting from 7 h after the conclusion of implantation. This time was necessary to allow safe transfer of the freshly implanted Y foil to the low-background chamber, due to the shorter-living activity of other A/q=135 species which was implanted in addition to
 470 Cs. The counting station is equipped with a Canberra BE3825 detector, with energy range from 3 keV to 3 MeV. The detector head is housed in a lead castle with N_2 atmosphere to reduce the background counts.

The decrease of the two γ emission peaks of ^{135m}Cs can be clearly observed as a function of time. Interfering emission at 785 keV is also detected. This is
 475 attributed to the products of the ^{135}I β -decay: the 785 keV photon is produced by the fast de-excitation of the daughter nucleus ^{135}Xe [34]. The longer half-life ($T_{1/2}=6.58$ h) of ^{135}I allows one to discriminate and remove its contribution. By combining together data from the two γ peaks typical of the spectrum of ^{135m}Cs , we measured a half-life of $T_{1/2}^{\text{exp}}=(53.5 \pm 1.4)$ min. The uncertainty in this case
 480 is purely statistical. The measured half-life is consistent with the previously reported value $T_{1/2} = (53 \pm 2)$ min [35]. This confirms the identification of ^{135m}Cs .

6. Conclusions and Outlook

We have designed, built, and tested an experimental facility for laser cooling
 485 and trapping of Cs isotopes and isomers. Production of the desired radioactive species is achieved either by proton-induced fission of U targets, or by a fusion-evaporation reaction. Nuclei are extracted as +1 ions from the target area, electrostatically accelerated, and mass separated in the IGISOL facility. Ions are implanted in a thin Y foil, where they dissipate their kinetic energy,

acquire an extra electron (“neutralization”), and are stored before undergoing thermal diffusion to the experimental chamber. Here, a PDMS organic coating allows efficient laser trapping. Tests conducted with a local beam of $^{133}\text{Cs}^+$ confirmed the correct operation of the system, and highlighted the direct contribution of the ion beam to the population of a magneto-optical trap. Atoms are cooled down to 150 μK , with a density of 10^{10} cm^{-3} . Preliminary estimates with $^{133}\text{Cs}^+$ identified a lower limit for the trapping efficiency of $\eta \geq 3 \times 10^{-6}$. This is compatible with successful experiments with other radioactive species [27], with potential for further improvement [36]. This opens up new scenarios and possibilities, from high-precision laser spectroscopy of unknown electronic states for investigations on the isomeric charge radius anomaly, to nuclear forensics. The experiment also constitutes the building block of a long-term endeavor for multi-body physics in ultra-cold isomeric gases.

In the immediate future, the ^{133}Cs MOT will be investigated over a wider range of implantation rates, following the installation of new particle diagnostics in the beam line. This will allow us to extend testing to lower ionic currents and to measure the overall efficiency of the setup. Ultimately, this will identify the lower trapping limit as a function of the number of implanted ions. This information is paramount to tailor the most effective approach for laser cooling and trapping of radioactive species.

Acknowledgements

This work was partially funded by the H2020-EU.1.3.2 programme through the Marie Curie Fellowship 2020-MSCA-IF-2014 “GAMMALAS” to L. M. (Proj. Ref. 657188), and by the Royal Society (IE151199). P. M. W. acknowledges STFC support under grant no. ST/L005743/1.

We thank Krishna Jadeja (UCL) and Yuval Cohen (UCL), who recently joined this project after completion of the setup.

Declarations of Interest

Declaration of interest: none.

References

- 520 [1] H.-J. Kluge, W. Nörtershäuser, Lasers for nuclear physics, *Spectrochimica Acta Part B: Atomic Spectroscopy* 58 (6) (2003) 1031 – 1045, IAP-2002 INTERNATIONAL CONFERENCE ON LASER PROBING.
doi:[https://doi.org/10.1016/S0584-8547\(03\)00063-6](https://doi.org/10.1016/S0584-8547(03)00063-6).
URL <http://www.sciencedirect.com/science/article/pii/S0584854703000636>
525
- [2] A. Pálffy, Nuclear effects in atomic transitions, *Contemporary Physics* 51 (6) (2010) 471–496. arXiv:<https://doi.org/10.1080/00107514.2010.493325>, doi:10.1080/00107514.2010.493325.
URL <https://doi.org/10.1080/00107514.2010.493325>
- 530 [3] B. Cheal, K. T. Flanagan, Progress in laser spectroscopy at radioactive ion beam facilities, *Journal of Physics G: Nuclear and Particle Physics* 37 (11) (2010) 113101.
URL <http://stacks.iop.org/0954-3899/37/i=11/a=113101>
- [4] K. Blaum, J. Dilling, W. Nörtershäuser, Precision atomic physics techniques for nuclear physics with radioactive beams, *Physica Scripta* 2013 (T152) (2013) 014017.
535 URL <http://stacks.iop.org/1402-4896/2013/i=T152/a=014017>
- [5] P. Campbell, I. Moore, M. Pearson, Laser spectroscopy for nuclear structure physics, *Progress in Particle and Nuclear Physics* 86 (2016) 127 – 180. doi:<https://doi.org/10.1016/j.ppnp.2015.09.003>.
540 URL <http://www.sciencedirect.com/science/article/pii/S0146641015000915>

- [6] J. A. Behr, G. Gwinner, Standard model tests with trapped radioactive atoms, *Journal of Physics G: Nuclear and Particle Physics* 36 (3) (2009) 033101.
 545 URL <http://stacks.iop.org/0954-3899/36/i=3/a=033101>
- [7] L. Willmann, K. Jungmann, C. J. G. Onderwater, R. G. E. Timmermans, H. W. Wilschut, Trapped radioactive isotopes for fundamental symmetry investigations, *Hyperfine Interactions* 211 (1) (2012) 39–43. doi:10.1007/s10751-012-0603-x.
 550 URL <https://doi.org/10.1007/s10751-012-0603-x>
- [8] S. Agustsson, G. Bianchi, R. Calabrese, L. Corradi, A. Dainelli, A. Khanbekyan, C. Marinelli, E. Mariotti, L. Marmugi, G. Mazzocca, L. Moi, L. Ricci, L. Stiazzini, L. Tomassetti, Observation of $7p^2P_{3/2} \rightarrow 7d^2D$ optical transitions in 209 and 210 francium isotopes, *Opt. Lett.* 42 (18) (2017) 3682–3685. doi:10.1364/OL.42.003682.
 555 URL <http://ol.osa.org/abstract.cfm?URI=ol-42-18-3682>
- [9] M. R. Kalita, J. A. Behr, A. Gorelov, M. R. Pearson, A. C. DeHart, G. Gwinner, M. J. Kossin, L. A. Orozco, S. Aubin, E. Gomez, M. S. Safronova, V. A. Dzuba, V. V. Flambaum, Isotope shifts in the $7s \rightarrow 8s$ transition of francium: Measurements and comparison to ab initio theory, *Phys. Rev. A* 97 (2018) 042507. doi:10.1103/PhysRevA.97.042507.
 560 URL <https://link.aps.org/doi/10.1103/PhysRevA.97.042507>
- [10] G. Gwinner, E. Gomez, L. A. Orozco, A. Perez Galvan, D. Sheng, Y. Zhao, G. D. Sprouse, J. A. Behr, K. P. Jackson, M. R. Pearson, S. Aubin, V. V. Flambaum, Fundamental symmetries studies with cold trapped francium atoms at isac, *Hyperfine Interactions* 172 (1) (2006) 45–51. doi:10.1007/s10751-007-9523-6.
 565 URL <https://doi.org/10.1007/s10751-007-9523-6>
- [11] Y. Sakemi, K. Harada, T. Hayamizu, M. Itoh, H. Kawamura, S. Liu, H. S. Nataraj, A. Oikawa, M. Saito, T. Sato, H. P. Yoshida, T. Aoki,
 570

- A. Hatakeyama, T. Murakami, K. Imai, K. Hatanaka, T. Wakasa, Y. Shimizu, M. Uchida, Search for a permanent edm using laser cooled radioactive atom, *Journal of Physics: Conference Series* 302 (1) (2011) 012051.
 URL <http://stacks.iop.org/1742-6596/302/i=1/a=012051>
- [12] E. Mariotti, A. Khanbekyan, C. Marinelli, L. Marmugi, L. Moi, L. Corradi, A. Dainelli, R. Calabrese, G. Mazzocca, L. Tomasseti, Francium trapping at the infn-lnl facility, *International Journal of Modern Physics E* 23 (05) (2014) 1430009. arXiv:<https://www.worldscientific.com/doi/pdf/10.1142/S0218301314300094>, doi:10.1142/S0218301314300094.
 URL <https://www.worldscientific.com/doi/abs/10.1142/S0218301314300094>
- [13] M. Bissell, K. Flanagan, M. Gardner, M. Avgoulea, J. Billowes, P. Campbell, B. Cheal, T. Eronen, D. Forest, J. Huikari, A. Jokinen, I. Moore, A. Nieminen, H. Penttilä, S. Rinta-Antila, B. Tordoff, G. Tungate, J. Äystö, On the decrease in charge radii of multi-quasi particle isomers, *Physics Letters B* 645 (4) (2007) 330–334. doi:<https://doi.org/10.1016/j.physletb.2006.12.054>.
- [14] G. D. Dracoulis, P. M. Walker, F. G. Kondev, Review of metastable states in heavy nuclei, *Reports on Progress in Physics* 79 (7) (2016) 076301.
 URL <http://stacks.iop.org/0034-4885/79/i=7/a=076301>
- [15] M. Di Rosa, S. Crane, J. Kitten, W. Taylor, D. Vieira, X. Zhao, Magneto-optical trap and mass-separator system for the ultra-sensitive detection of ^{135}Cs and ^{137}Cs , *Applied Physics B* 76 (1) (2003) 45–55. doi:10.1007/s00340-002-1063-9.
 URL <https://doi.org/10.1007/s00340-002-1063-9>
- [16] G. Yang, H. Tazoe, M. Yamada, ^{135}Cs activity and $^{135}\text{Cs}/^{137}\text{Cs}$ atom ratio in environmental samples before and after the fukushima daiichi nuclear

power plant accident, Scientific Reports 6 (2016) 24119 EP –.

URL <http://dx.doi.org/10.1038/srep24119>

- [17] L. Marmugi, P. M. Walker, F. Renzoni, Coherent gamma photon generation in a bose–einstein condensate of 135mcs, Physics Letters B 777 (2018) 281 – 285. doi:<https://doi.org/10.1016/j.physletb.2017.12.036>.

URL <http://www.sciencedirect.com/science/article/pii/S0370269317310134>

- [18] J. Persson, Table of hyperfine anomaly in atomic systems, Atomic Data and Nuclear Data Tables 99 (1) (2013) 62 – 68. doi:<https://doi.org/10.1016/j.adt.2012.04.002>.

URL <http://www.sciencedirect.com/science/article/pii/S0092640X1200085X>

- [19] I. Moore, T. Eronen, D. Gorelov, J. Hakala, A. Jokinen, A. Kankainen, V. Kolhinen, J. Koponen, H. Penttilä, I. Pohjalainen, M. Reponen, J. Rissanen, A. Saastamoinen, S. Rinta-Antila, V. Sonnenschein, J. Äystö, Towards commissioning the new igisol-4 facility, Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms 317 (2013) 208 – 213, xVith International Conference on ElectroMagnetic Isotope Separators and Techniques Related to their Applications, December 2–7, 2012 at Matsue, Japan. doi:<https://doi.org/10.1016/j.nimb.2013.06.036>.

URL <http://www.sciencedirect.com/science/article/pii/S0168583X13007143>

- [20] Al-Adili, A., Jansson, K., Lantz, M., Solders, A., Gorelov, D., Gustavsson, C., Mattera, A., Moore, I., Prokofiev, A. V., Rakopoulos, V., Penttilä, H., Tarrío, D., Wiberg, S., Österlund, M., Pomp, S., Simulations of the fission-product stopping efficiency in igisol, Eur. Phys. J. A 51 (5) (2015) 59. doi:[10.1140/epja/i2015-15059-2](https://doi.org/10.1140/epja/i2015-15059-2).

URL <https://doi.org/10.1140/epja/i2015-15059-2>

- [21] P. Karvonen, I. Moore, T. Sonoda, T. Kessler, H. Penttilä, K. Peräjärvi, P. Ronkanen, J. Äystö, A sextupole ion beam guide to improve the efficiency and beam quality at igisol, Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms 266 (21) (2008) 4794 – 4807. doi:<https://doi.org/10.1016/j.nimb.2008.07.022>. URL <http://www.sciencedirect.com/science/article/pii/S0168583X08009191>
- [22] L. J. Vormawah, M. Vilén, R. Beerwerth, P. Campbell, B. Cheal, A. Dicker, T. Eronen, S. Fritzsche, S. Geldhof, A. Jokinen, S. Kelly, I. D. Moore, M. Reponen, S. Rinta-Antila, S. O. Stock, A. Voss, Isotope shifts from collinear laser spectroscopy of doubly charged yttrium isotopes, Phys. Rev. A 97 (2018) 042504. doi:[10.1103/PhysRevA.97.042504](https://doi.org/10.1103/PhysRevA.97.042504). URL <https://link.aps.org/doi/10.1103/PhysRevA.97.042504>
- [23] R. Nolli, M. Venturelli, L. Marmugi, A. Wickenbrock, F. Renzoni, Compact setup for the production of ^{87}Rb $|f = 2, m_f = +2\rangle$ bose-einstein condensates in a hybrid trap, Review of Scientific Instruments 87 (8) (2016) 083102. arXiv:<https://doi.org/10.1063/1.4960395>, doi:[10.1063/1.4960395](https://doi.org/10.1063/1.4960395). URL <https://doi.org/10.1063/1.4960395>
- [24] M. Stephens, C. Wieman, High collection efficiency in a laser trap, Phys. Rev. Lett. 72 (1994) 3787–3790. doi:[10.1103/PhysRevLett.72.3787](https://doi.org/10.1103/PhysRevLett.72.3787). URL <https://link.aps.org/doi/10.1103/PhysRevLett.72.3787>
- [25] Z.-T. Lu, K. L. Corwin, K. R. Vogel, C. E. Wieman, T. P. Dinneen, J. Maddi, H. Gould, Efficient collection of ^{221}Fr into a vapor cell magneto-optical trap, Phys. Rev. Lett. 79 (1997) 994–997. doi:[10.1103/PhysRevLett.79.994](https://doi.org/10.1103/PhysRevLett.79.994). URL <https://link.aps.org/doi/10.1103/PhysRevLett.79.994>
- [26] S. Agustsson, G. Bianchi, R. Calabrese, L. Corradi, A. Dainelli, A. Khanbekyan, C. Marinelli, E. Mariotti, L. Marmugi, L. Ricci, L. Stiazzini,

- L. Tomassetti, A. Vanella, Enhanced atomic desorption of 209 and 210
 660 francium from organic coating, Scientific Reports 7 (1) (2017) 4207. doi:
 10.1038/s41598-017-04397-y.
 URL <https://doi.org/10.1038/s41598-017-04397-y>
- [27] G. Gwinner, J. A. Behr, S. B. Cahn, A. Ghosh, L. A. Orozco, G. D. Sprouse,
 F. Xu, Magneto-optic trapping of radioactive ^{79}Rb , Phys. Rev. Lett. 72
 665 (1994) 3795–3798. doi:10.1103/PhysRevLett.72.3795.
 URL <https://link.aps.org/doi/10.1103/PhysRevLett.72.3795>
- [28] J. F. Ziegler, M. Ziegler, J. Biersack, Srim – the stopping and range
 of ions in matter (2010), Nuclear Instruments and Methods in Physics
 Research Section B: Beam Interactions with Materials and Atoms 268 (11)
 670 (2010) 1818 – 1823, 19th International Conference on Ion Beam Analysis.
 doi:<https://doi.org/10.1016/j.nimb.2010.02.091>.
 URL <http://www.sciencedirect.com/science/article/pii/S0168583X10001862>
- [29] K.-H. Weber, C. J. Sansonetti, Accurate energies of ns, np, nd, nf, and ng
 675 levels of neutral cesium, Phys. Rev. A 35 (1987) 4650–4660. doi:10.1103/
 PhysRevA.35.4650.
 URL <https://link.aps.org/doi/10.1103/PhysRevA.35.4650>
- [30] V. Coppolaro, N. Papi, A. Khanbekyan, C. Marinelli, E. Mariotti, L. Mar-
 mugi, L. Moi, L. Corradi, A. Dainelli, H. Arikawa, T. Ishikawa, Y. Sakemi,
 680 R. Calabrese, G. Mazzocca, L. Tomassetti, L. Ricci, Light desorption from
 an yttrium neutralizer for rb and fr magneto-optical trap loading, The Jour-
 nal of Chemical Physics 141 (13) (2014) 134201. doi:10.1063/1.4896609.
- [31] A. Gorelov, J. A. Behr, M. R. Kalita, M. R. Pearson, M. Tandecki, D. L,
 S. Aubin, R. Collister, A. C. Dehart, G. Gwinner, E. Gomez, L. A. Orozco,
 685 J. Z, Neutralizer for triumf’s experiment for measurements of parity non-
 conservation in francium, in: 2016 Fall Meeting of the APS Division of

Nuclear Physics, Vol. 61, 2016, abstract ID: BAPS.2016.DNP.HG.7.

URL <http://meetings.aps.org/link/BAPS.2016.DNP.HG.7>

- [32] K. L. Corwin, Z.-T. Lu, C. F. Hand, R. J. Epstein, C. E. Wieman,
 690 Frequency-stabilized diode laser with the zeeman shift in an atomic vapor, Appl. Opt. 37 (15) (1998) 3295–3298. doi:10.1364/AO.37.003295.
 URL <http://ao.osa.org/abstract.cfm?URI=ao-37-15-3295>

- [33] L. Marmugi, E. Mariotti, A. Burchianti, S. Veronesi, L. Moi, C. Marinelli,
 695 Laser-driven self-assembly of shape-controlled potassium nanoparticles in porous glass, Laser Physics Letters 11 (8) (2014) 085902.
 URL <http://stacks.iop.org/1612-202X/11/i=8/a=085902>

- [34] B. Singh, A. A. Rodionov, Y. L. Khazov, Nuclear data sheets for $a = 135$, Nuclear Data Sheets 109 (3) (2008) 517 – 698.
 doi:<https://doi.org/10.1016/j.nds.2008.02.001>.
 700 URL <http://www.sciencedirect.com/science/article/pii/S0090375208000094>

- [35] H. Warhanek, A 53 min isomer of caesium 135, Nuclear Physics 33 (1962) 639 – 647. doi:[https://doi.org/10.1016/0029-5582\(62\)90555-2](https://doi.org/10.1016/0029-5582(62)90555-2).
 705 URL <http://www.sciencedirect.com/science/article/pii/0029558262905552>

- [36] S. Aubin, E. Gomez, L. A. Orozco, G. D. Sprouse, High efficiency magneto-optical trap for unstable isotopes, Review of Scientific Instruments 74 (10) (2003) 4342–4351. arXiv:<https://doi.org/10.1063/1.1606093>, doi:10.1063/1.1606093.
 710 URL <https://doi.org/10.1063/1.1606093>